


Estimating Mold Spores by IAQ Pro Sample Value (SV)*	
SV	Total Spore Estimate
< 0.012	Not Detected
0.012	3,000 - 10,000
0.040	15,000 - 40,000
> 0.100	> 75,000



* Sample Values represent the relative intensity of the Sample line as compared to the Control line. As different species of molds react differently with the IAQ Pro Asp/Pen test, ONLY relative ranges of spore levels can be estimated from the SV test result. If precise spore levels are desired, samples should be submitted to a qualified reference laboratory.